



**REGIONAL INSTITUTE OF MEDICAL SCIENCES: IMPHAL-795004**  
(An Autonomous Institute under Ministry of Health & Family Welfare, Govt. of India.)

No. 75-Project/Pead-RIMS-14

Imphal the **28 MAY 2015**

To

Subject:- Submission of Rate Quotations along with Dealership/Authorized Stockiest /Manufacturer, Certificate, Terms and Conditions for supply and installation of 1 (one) no. each of (i) Deep freezer -80°C and (ii) deep freezer -20°C.

Dear Sirs,

I am directed to request you kindly to submit the rate quotation along with a copy of dealership /authorized stockiest/manufacturer certificate, terms and conditions for the following items on or before 05/06/2015 and bids will be opened on 06/06/2015 at 2:00 p.m.

“(i) Deep freezer -80°C & (ii) deep freezer -20°C” in favor of the Director RIMS, Imphal

**The following details, terms and Conditions may please be furnished.**

- (1) The price in words and figures should be indicated clearly.
- (2) The warranty period of the equipments is 3 (three) years. If not mentioned, it will be assured that the warranty is for 3 years.
- (3) The warranty period of the equipments with AMC & CAMC charges for next 3 years should be mentioned.
- (4) The cover of the rate quotation should be subscribed as “(i) Deep freezer -80°C and (ii) deep freezer -20°C. (Specifications at annexure-A) for use of Pediatrics Department, RIMS, Imphal.
- (5) VAT charges for the items should be mention separately.
- (6) The rate quoted should be valid for one year from the due date of submitting the tenders.
- (7) Firm's TIN No., Registration No., PAN No., etc should be furnished. PAN card should be in the name of the firm or proprietor's name.
- (8) The supply should be made within 4 weeks from the date of placing supply order.
- (9) Failure to supply the items within the stipulated period and also within the validity period, the successful bidder/supplier is liable to be blacklisted.
- (10) Earnest money Rs.10,000/- (Rupees ten thousand) only along with rate quotation should be submitted in the form of Bank Draft from any Nationalized/Schedule Bank in favor of Director RIMS, payable at Imphal. If earnest money is not submitted, rate quotations shall be rejected summarily. Those firms who already submit EMD at previous rate quotations are not required to submit again.
- (11) Product catalogues should be enclosed along with the quotations.
- (12) Client list should be enclosed in the quotations.

Yours faithfully


(Y. Rajendra Singh)

Deputy Director (Admin),

Regional Institute of Medical Sciences,  
Imphal.

## Annexure - A

Sl No	Name of the instrument	Specifications	Qty	Accessories
1	<b>Deep Freezer (-80 °C)</b>	<ol style="list-style-type: none"> <li>1. Capacity of 400- 410 liters; 14.5 Cu. Ft (approx) - preferably upright.</li> <li>2. Programmable temperature range from -50°C to -86°C in 1°C increments even at ambient temperature upto 32°C, suitable for preservation of bio-medical products , bacteria, virus etc.</li> <li>3. Microprocessor-controlled temperature facility and alarms with non-volatile memory.</li> <li>4. Alarms: Audible and Visible alarms for Temperature filter clean, power out, Low Battery, System Fail, and Fault analysis. Battery Backup: Activates alarms and display temperature during power outage.</li> <li>5. Freezer should have Vent Plunger to prevent vacuum formation, Keyed lock, polished 304Lstainless steel interior, exterior : @ 18 gauge steel, 1.2 mm thick (SS -430 grade) powder coated, scratch and rust resistant</li> <li>6. Micro Processor control of temperature and alarms with non-volatile memory</li> <li>7. Five compartment with 4 adjustable height corrosion resistant Stainless Steel Shelves. Each compartment is well equipped with Insulated Inner Door with magnetic lock.</li> <li>8. Insulated with Polyurethane foam, approx.5.1", 130 mm thick</li> <li>9. Hermetically-sealed two stage cascade system with capacity to cope in high-ambient conditions</li> <li>10. Minimum Power Consumption approx. 10-12.6 kWh/Day(power supply 220-230 V/50Hz)</li> <li>11. Facility for Automatic Reset, Automatic Restart with Non-Volatile memory, programmed Restart.</li> <li>12. Hermetically-sealed two-stage cascade refrigeration system with CFC, HCFC free refrigerants.</li> <li>13. Internal Dimensions 126.5 cm x 55 cm x 57.5 cm (H x W x D) approx, External: 191.5 cm x 80 cm x 85.2 cm (H x W x D) approx.</li> <li>14. Net Weight 235 Kg (approx). and Shipping Weight 283 Kg (approx).</li> <li>15. Unique 4-digit Password to prevent unauthorized changes(preferred).</li> <li>16. S.M.A.R.T. Plus™ Diagnostic Software to trace and solve system errors.</li> <li>17. CE Certified and UL Certified</li> </ol>	1	<ol style="list-style-type: none"> <li>1. Power back –up in event of power failure for continuous display of chamber temperature</li> <li>2. Servo stabilizer (5KVA)</li> </ol>



(Y. Rajendra Singh)

Deputy Director (Admn)

Regional Institute of Medical Sciences,  
Imphal

## Annexure - A


Sl No	Name of the instrument	Specifications	Qty	Accessories
2	<b>Deep Freezer (-20 °C)</b>	<ol style="list-style-type: none"> <li>1. Vertical type;</li> <li>2. Temp: -20 Deg C (Accuracy <math>\pm 1^{\circ}\text{C}</math>)</li> <li>3. With voltage stabilizer</li> <li>4. Built in condenser for low energy consumption</li> <li>5. No dust, low noise</li> <li>6. Combination of pull out drawers and fast freeze shelves</li> <li>7. Control panel at the top of the cabinet with thermometer or temperature warning light</li> <li>8. Thermostat</li> <li>9. Fast freeze button and lamp</li> <li>10. Mains warning lamp</li> <li>11. Single door</li> <li>12. CFC free heavy duty compressors</li> <li>13. Provided with 8 shelves (approx)</li> <li>14. Gross capacity 345 liters</li> <li>15. Net capacity 250 - 300 liters</li> <li>16. Power supply – 220-230 V</li> </ol>	<b>1</b>	1. Servo stabilizer (5KVA)

### Terms and Conditions

- a. Price for all instruments should be up to installation at Regional Institute of Medical Sciences, Imphal. Charges for clearing and transportation should be incurred by the supplier.
- b. 24 months warranty from the date of installation.
- c. Laboratory demonstration at the time of installation.
- d. All prices should be quoted in Indian rupees (INR).

### Desirable:

- a. CMC next two years



(Y. Rajendra Singh)

Deputy Director (Admn)

Regional Institute Of Medical Sciences,  
Imphal