



Point no I of technical specification of CBCT with turnkey work should read as "The CBCT should possess field of view (FOV) flexibility through the articulated robotic arm positioning. Field of View (FOV) - Flexible in the range of ~~40x40~~ 40mm x 40mm or smaller upto 150mm x 120mm or larger."


K. Sathikumar Singh
(DR. KOIJAM SATHIKUMAR SINGH)


(DR. ALBERT ASHEM)

SD (P)
RIMS

Deepak N Singh
(DR. DEEPAK NINGOMBAM SINGH)


(DR. B. PRAVEEN REDDY)


1/2/25